

Programing a New Infusion



1. Select "New Infusion"



2. Select "Program"



3. Select "TPN Mode"



4. Input Time and Volume

FOR INTERMITTENT INFUSION ONLY



5. Select "Ramp"



6. Input Up/Down Time